



| CAT. NO. | METAL | CERVICAL PLATE |
|----------|-------|-----------------------|
| 690.122 | TT | Cervical Plate 22.5mm |
| 690.125 | TT | Cervical Plate 25.0mm |
| 690.127 | TT | Cervical Plate 27.5mm |
| 690.130 | TT | Cervical Plate 30.0mm |
| 690.132 | TT | Cervical Plate 32.5mm |
| 690.135 | TT | Cervical Plate 35.0mm |
| 690.137 | TT | Cervical Plate 37.5mm |
| 690.140 | TT | Cervical Plate 40.0mm |
| 690.142 | TT | Cervical Plate 42.5mm |
| 690.145 | TT | Cervical Plate 45.0mm |
| 690.147 | TT | Cervical Plate 47.5mm |
| 690.150 | TT | Cervical Plate 50.0mm |
| 690.152 | TT | Cervical Plate 52.5mm |
| 690.155 | TT | Cervical Plate 55.0mm |
| 690.157 | TT | Cervical Plate 57.5mm |
| 690.160 | TT | Cervical Plate 60.0mm |
| 690.162 | TT | Cervical Plate 62.5mm |
| 690.165 | TT | Cervical Plate 65.0mm |
| 690.167 | TT | Cervical Plate 67.5mm |
| 690.170 | TT | Cervical Plate 70.0mm |

Specification

- Cervical Plates are available in Titanium.
- Available length 22.5-70mm.

Indication for use

- The Anterior Cervical Spine Plate system is intended for anterior underbody screw fixation of cervical spine.
- The system is indicated for use in the temporary stabilization or the anterior spine during the development of cervical spinal fusions in patients with degenerative disc disease, trauma (including fractures), tumors, deformity (defined as kyphosis, lordosis, or scoliosis), Pseudoarthrosis, and/or failed previous fusions.

